

**PRODUCT NAME: Bialex LDL**

LOT 24LDL0902 DATE MANUFACTURER: 2024-09-03

EXP. DATE: 2026-08-30 REF. # GQ-LDL-3010

Qty: 100 Kits

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable

| Box: Acceptable

Method: Direct

Performance Characteristics:**Reagent 1:**

Aspect Liquid Translucid Light brown, non-visible particles, w/o surfactant

Reagent 2:

Aspect Translucid, non-visible particles.

Color reaction: Purple

	QCNet (Bio-Rad)	Results @ 600nm
Control 1	153.0 mg/dL	154.0 mg/dL (n=20)
Control 2	52.3 mg/dL	52.0 mg/dL (n=20)
Comparable method with Beckman Coulter AU System		
PassCheck -Lot 24PCK0401 (24PCK0814)		
Control 1	54 (56 mg/dL)	
Control 2	130 (132 mg/dL)	
Serum Abnormal >0.160	Pass	
Serum Normal <0.090	Pass	

PassCheck CAL Ref: GQ-LIP-1, Lot: 24LIP0901

Value: 45.48 mg/dL

Conclusion: Release Batch**Liabe Person:**

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.**TollFree: 833-418-2116 email: info@bialexlab.com**