



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Creatinine

LOT 24CRE1201 **DATE MANUFACTURER:** 2024.12.02

EXP. DATE: 2026-09-30 **REF. #** GQ-CRE-125

QTY: 266 kits

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable

Performance Characteristics:

Reagent Creatinine:

Aspect Liquid Translucid Yellow, non-visible particles, with surfactant

Standard: Aspect Liquid Translucid, nonvisible particles, w/o surfactant,
concentration **5mg/dL**

Reagent	▲ Kinetic @37°C	Results @505nm
Reagent Creatinine	Vx 105/1min	0.176
Creatinine	QCNet (Bio-Rad) Lot 89740	Results @ 505nm
Control 2	n=20 / 6.15 mg/dL	6.00 (+/- 0.1%)
Control 1	n=20 / 2.47 mg/dL	2.60 (+/- 1.0%)
Comparable method with Abbott Infinity		

PassCheck Control Level 1 and 2 - Lot 24PCK0814 (Ref: GQ-PC1-1, GQ-PC2-1)

Control 1 0.94 mg/dL (1.10 mg/dL)

Control 2 4.00 mg/dL (4.30 mg/dL)

PassCheck CAL-Lot 24CAL0701

Value: Pending

Storage: 2-8°C

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®