

**PRODUCT NAME: Bialex LDL**

LOT 25LDL0101 DATE MANUFACTURER: 2025-01-10

EXP. DATE: 2027-01-30 REF. # GQ-LDL-3010

Qty: 20 Kits

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable| HPDE Bottle: Acceptable

| Box: Acceptable

Method: Direct

Performance Characteristics:**Reagent 1:**

Aspect Liquid Translucid Light brown, non-visible particles, w/o surfactant

Reagent 2:

Aspect Translucid, non-visible particles.

Color reaction: Purple

	QCNet (Bio-Rad)	Results @ 600nm
Control 1	153.0 mg/dL	150.0 mg/dL (n=20)
Control 2	52.3 mg/dL	51.0 mg/dL (n=20)

Comparable method with Beckman Coulter AU System

PassCheck -Lot 24PCK0121

Control 1	Pending mg/dL
Control 2	Pending mg/dL:
Serum Abnormal	Pass
Serum Normal	Pass

PassCheck CAL Ref: GQ-LIP-1, Lot: 24LIP0901

Value: Pending

Conclusion: Release Batch**Liabe Person:**

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.**TollFree: 833-418-2116 email: info@bialexlab.com**