



Certificate of Analysis

Date: 2025.02.01
Lot: 25K0201
Property Name: BIALEX K

Batch: 20 Kits

Reference: GQ-K-100
Use: IN VITRO DIAGNOSTIC
Expiration Date: 2027-01-01 (YYYY-MM-DD)

Characteristics:
REAGENT AND STANDARD BETWEEN RANGE CHARACTERISTIC
Control Reference:

RANGE VALUES
CONTROL 1: 4.6
CONTROL 2 6.2
Units: mEq/L
St. Value: 4mEq/L

Result: LEVEL 1 4.1 / LEVEL 2 6.0 mEq/L

Conclusion: Release for sale

[Handwritten signature]

A circular stamp with a blue border. Inside, the letters "QC" are prominently displayed in the center. Above "QC" is the word "BIALEX" and below it is "QUALITY ASSURANCE". There is also a small "DATE" label next to the "QC" text.