



CERTIFICATE OF ANALYSIS

Date: 2025-02-20 (YYYY-MM-DD)
Lot: 25NA0201
Batch: 20 Kits

Property Name: BIALEX Na
Ref: GQ-NA-100
Use: IN VITRO DIAGNOSTIC
Expiration Date: 2027-02-15

Appearance: Color, clarity between parameters

Chemical Stability: Pass

Handling Stress: Pass

Accelerated stability: Pass

Control Reference Values

MEDIA VALUES:

CONTROL 1: 128.50 / CONTROL 2: 150

Unites: mEqL

Result:

LEVEL 1: 130 mEqL / LEVEL 2: 162 mEqL

This lot is ready for release: Ready



J. TANAKA