

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Dyazo Direct Bilirrubin

LOT 25BD0301 DATE MANUFACTURER: 2025-03-14

EXP. DATE: 2027-03-10 (YYYY-MM-DD) REF. # GQ-BD-100, GQ-BD-125

LOT QTY: 10L

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Total Bilirrubin

Aspect Liquid Translucid, nonvisible particles, with surfactant

Reagent Nitrite:

Aspect Liquid Translucid, nonvisible particles, with surfactant

Reagent	Values @ 550nm	Results @ 550nm
Abs Work Reagent	(0.038-0.200)	0.200 (after 8 hours)

D.Bilirrubin	QCNet (Bio-Rad) Lot 89740	Results @ 550nm
Control 1	n=20 / 0.379 mg/dL	0.360 (+/- 0.1%)
Control 2	n=20 / 1.680 mg/dL	1.710 (+/- 0.1%)
Comparable result with Beckman Coulter Au Systems		

GQ-PC-1 PassCheck Control Lot 25PCK0131 & 25PCK0132

Control 1	Control 2
0.50 mg/dL	4.70 mg/dL
(0.52 mg/dL)	(4.73mg/dL)

Time reader for Abs:	5 minutes	8 minutes
----------------------	-----------	-----------

FACTOR: 11.45

Liabile Person:

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 0.46 mg/dL

Storage: 2-25°C

