

Certificate of Analysis

Product Name	Sodium (Na) Test Kit (Enzymatic Method)			Material Code	71339
Lot No.	50220151N	Package	R1: 18mL×4 R2: 6mL×4;	Manufacture Date	2025.2.22
Lot Size	R1 : 18mL×4 R2 : 6mL×4;	Market Area	Bialex Formulation	Expiry Date	2026.2.21
Sample Amount	1盒	Sampling Date	2025.2.22	Analysis Date	2025.2.22
Inspection Basis	Inspection procedures				
Test Items	Quality Standard		Result		Comment
Appearance and Character					Qualified
	Reagent componets are complete.Labels are clear and correct.		Reagent componets are complete.Labels are clear and correct.		
	R1 is colorless clear liquid		R1 is colorless clear liquid		
	R2 is colorless clarifying liquid		R2 is colorless liquid		
Volume	R1≥18 mL, R2≥6 mL R1≥18 mL, R2≥6 mL		R1 18.4 ml R2 6.2ml		符合标准
空白吸光度 Blank Absorbance	Reagent blank absorbance≤0.5,(405nm,lcm optical path)		0.2363		Qualified
Accuracy	Whether the quality control results are within the target range		Yes		
Linear Range	range: [80,180]mmol/L,r≥0.990		0.9990		Qualified
	repeatability CV≤5%		1.46%		Qualified
Conclusion	Pass				
Remarks					
QC		Checked By	JIM	Approved By	D. KLEIN
Signature / Date	2025.2.22	Signature / Date	2025.2.22	Signature / Date	2025.2.22