



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Creatinine

LOT 25CRE0301 **DATE MANUFACTURER:** 2025.03.04

EXP. DATE: 2027.03.10 (YYYY.MM.DD) **REF. #** GQ-CRE-125

QTY: 120L

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable

Performance Characteristics:

Reagent Creatinine:

Aspect Liquid Translucid Yellow, non-visible particles, with surfactant

Standard: Aspect Liquid Translucid, nonvisible particles, w/o surfactant,
concentration **4.5mg/dL**

Reagent
Reagent Creatinine

▲ Kinetic @37°C
Vx 105/1min

Abs Reagent @505nm
0.174

Creatinine

QCNet (Bio-Rad) Lot 89740

Results @ 505nm

Control 2

n=20 / 6.15 mg/dL

5.98 (+/- 0.1%)

Control 1

n=20 / 2.47 mg/dL

2.40 (+/- 1.0%)

Comparable method with Abbott Infinity

PassCheck Control Level 1 and 2 - 25PCK0121 / 24PCK0121

Control 1

0.96 mg/dL (1.00 mg/dL)

Control 2

4.50 mg/dL (4.48 mg/dL)

PassCheck CAL-Lot 24CAL0701

Value: 1.02

Storage: 2-8°C

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®