



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Precipitation HDL
LOT 25HDP0301 **DATE MANUFACTURER:** 2023-03-03
EXP. DATE: 2027-03-10 (YYYY-MM-DD), **REF. #** GQ-HDP-50, GQ-HDP-100

Qty Manufacture: 30 kits

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Precipitation HDL:

Aspect Liquid Translucid clear, non-visible particles, w/o surfactant

Standard: No include in the kit – Use Standard of Cholesterol Bialex (GQ-CC-1)

Reagent	Values @ 500nm	Results @ 500nm
Reagent HDL Precipitant	(0.010-0.900)	0.041

HDL Precipitant	QCNet (Bio-Rad) Lot 89740	Results @ 500nm
Control 1	n=20 / 57.6 mg/dL	59.72 mg/dL
Control 2	n=20 / 26.5 mg/dL	26.94 mg/dL

Comparable method with Siemens Dimensions Series.

PassCheck -Lot 25PCK0130

Control 1

Pending

Control 2

Pending

PassCheck CAL-Lot
Value: NRD

Time reader for Abs:	30 minutes	Pass
----------------------	------------	------

Storage: 2-25°C

Conclusion: Release for sale

Liabile Person:

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex[®]