

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex AST

BOX LOT 24AST1101/ **INNER LOT** 24AST110 **DATE MANUFACTURER:** 2024-11-11

EXP. DATE: 2026-11-01 **REF. #** GQ-AST-100

Qty Manufacture: 206 Kits

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent AST:

Aspect Liquid Translucid clear, non-visible particles.

Reagent

Reagent AST

Control Liquichek Lot: 92970

Control 1: 34 U/L

Results:

Control 1: 32 U/L

Reagent AST (AA)

Control 1

Control 2

Values @ 340nm

>900

Control 2: 190 U/L

Control 2: 180 U/L

PASSCHECK Results @ 340nm, Constant Tempeture @ 37°C

n=20 / 59 U/L

n=20 / 230 U/L

Results @ 340nm

1.800

Run time: 180s, Interval time: 60s, Leg time: 60s

Factor: 5185 (Using Biolis 24i)

Factor: 2000 (Most Instruments, Run time: 180s, Lag time: 60s)

Factor: 1768 (Cuvetta x 1cm)

PassCheck Calibrator, REF: GQ-CAL-5 -Lot 24PCK0801


Value: 54.26 U/L

Storage: 2-8°C

Conclusion: Release for sale

Liabe Person:




DROGUERIA
ALEMANA
Q.F. Elizabeth H. Pichardo Mendez
Química Farmacéutica
S. de C. 19907

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex®