

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Triglycerides

LOT 25TT041101 DATE MANUFACTURER: 2025-04-07 to 2025-04-13

EXP. DATE: 2026-11-30 (YYYY-MM-DD) REF. # GQ-TT-100, GQ-TT-125, GQ-TT-2X125, GQ-TT-4X125

Qty Manufacture: 5L

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable

Performance Characteristics:

Reagent Triglycerides:

Aspect Liquid Yellow (Source Candide Rugose) to pink,
non-visible particles, with surfactant 1%

Interference blocker

Standard:

Include in the kit – Use Standard of Triglycerides Bialex (GQ-TT-1),
Concentration 200mg/dL

Reagent

Values @ 505nm

Results @ 505nm

Reagent Triglycerides Abs:

QCNet (Bio-Rad) Lot 89740

Results @ 505nm

Reagent Triglycerides

Control 1

n=20 / 153 mg/dL

157.0 mg/dL

Control 2

n=20 / 77.8 mg/dL

74.1 mg/dL

Comparable method with Abbott Alinity c Analyzer (continued)

PassCheck -Lot 25PCK0131, 25PCK0132 (REF: GQ-PC1-1,GQ-PC2-1)

Control 1

100 mg/dL

Control 2

220 mg/dL

PassCheck CAL-Lot 24CAL0701 (REF: GQ-CAL-5)

Value 106 mg/dl

Recommendations: Ultra sensible to cross contamination, keep out of direct light.

Storage 2-8°C

Conclusion: Release for sale

Liable Person: Jean Tankaka

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex[®]

