

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Triglycerides

LOT 25TT0401 DATE MANUFACTURER: 2025-04-02 to 2025-04-07

EXP. DATE: 2027-04-10 (YYYY-MM-DD) REF. # GQ-TT-100, GQ-TT-4X125

Qty Manufacture: 40L

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable | Amber Bottle

Performance Characteristics:

Reagent Triglycerides:

Aspect Liquid Yellow (Source Candide Rugose) to pink,
visible particles, with surfactant >1%

Interference blocker - include

Standard:

Include in the kit – Use Standard of Triglycerides Bialex (GQ-TT-1),
Concentration 200mg/dL (do not exchange with another lot)

Reagent

Values @ 505nm

Results @ 505nm

Reagent Triglycerides Abs:

QCNet (Bio-Rad) Lot 89740

Results @ 505nm

Reagent Triglycerides

n=20 / 153 mg/dL

155.0 mg/dL

Control 1

n=20 / 77.8 mg/dL

74.10 mg/dL

Control 2

Comparable method with Abbott Alinity c Analyzer (continued)

PassCheck -Lot 25PCK0131, 25PCK0132 (REF: GQ-PC1-1,GQ-PC2-1)

Control 1

100 mg/dL

Control 2

220 mg/dL

PassCheck CAL-Lot 24CAL0701 (REF: GQ-CAL-5)

Value 103.86 mg/dl

Recommendations: Ultra sensible to cross contaminations, keep out of direct light.

Storage 2-8°C

Conclusion: Release for sale

Liable Person: Jean Tankaka

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex®

