

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Cholesterol Enzymatic Oxidase
LOT 25CHO0402 **DATE MANUFACTURER:** 2025-04-16 (YYYY-MM-DD)
EXP. DATE: 2027-03-15 **REF. #** GQ-CC-100

Qty: 10L

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Cholesterol:

Translucent, clear pink liquid. Contains visible enzyme microparticles that disappear with increasing temperature and the addition of sample.

Standard:

Include in the kit – Use Standard of Cholesterol Bialex (GQ-CC-1),
Concentration 200mg/dL

Abs@ 505nm (against DI Water) = 0.045

Reagent

Reagent Cholesterol

Reagent Cholesterol

Control 1

Control 2

Comparable method with Catachem ChemWell-T

Values @ 510nm

(0.010-0.900)

QCNet (Bio-Rad) Lot 89740

n=20 / 224 mg/dL

n=20 / 96 mg/dL

Results @ 505nm

0.045

Results @ 505nm

233.10 mg/dL

99.15 mg/dL

PassCheck Control (Ref: GQ-PC1-01, GQ-PC2-01) -Lot 25PCK0131, 25PCK0132 -@505nm:

Control 1

122.40 mg/dL

Control 2

265.00 mg/dL

PassCheck CAL-Lot 24CAL0701 (Ref: GQ-CA-5)

Value: 122.00 mg/dL

Time reader for Abs: 1 hour -Pass

Storage 2-8°C

Conclusion: Release for sale

Liabe Person

