



CERTIFICATE OF ANALYSIS

Product Name: Bialex Emulsion Lipase
Reference #: GQ-LL-60
Batch: 50 kits

LOT: 24LL0201 **Expiration Date:** 2026.02.01

Physical appearance: Substrate clear translucent: Good
Buffer clear Translucent: Good
Packaging Material: Good

Importante Notice: Don't use commercial controls for this method
Serum lipase activity should be validity by titrimetric method.

Work reagent: Pass
Factor: 4800 (AA instruments - Biolis)
Factor: in almost semi-automatic instruments: **11000 at 25°C**

The kit was evaluated with known 20 pathological and 10 normal samples and the result were in agreement

For validation use fresh valued serums

Conclusion: Release for sale

Handwritten signature