

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Total Protein
LOT 24TPB1101 **DATE MANUFACTURER:** 2024-11-25
EXP. DATE: 2026-11-01 **REF. #** GQ-TPB-125

Qty Manufacture: 80 Liters

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Total Protein:

Aspect Liquid Translucid-blue clear, non-visible particles, w/o surfactant

Standard:

Include in the kit – Use Standard of Bialex Total Protein (GQ-TPB-1),

Aspect Yellow

Concentration 5 g/dL

Reagent

Reagent Total Protein

Reagent Total Protein

Control 1

Control 2

Comparable method with Abaxis Piccolo.

Values @ 540nm

(0.060-0.900)

QCNet (Bio-Rad) Lot 89740

n=20 / 6.70 g/dL

n=20 / 4.30 g/dL

Results @ 540nm

0.081

Results @ 540nm

6.95 g/dL

4.35 g/dL

GQ-PC-1 PassCheck Multicontrol - LOT 24PCK0814 (24PCK0401)

Control 1

3.60 g/dL

Control 2

7.11 g/dL

Cal Value (24CAL0701) = 3.57 g/dL

Time reader for Abs: 45m

Pass

Storage: 2-8°C

Conclusion: Release for sale

Liable Person:

