

## CERTIFICATE OF ANALYSIS

**PRODUCT NAME:** Bialex Albumin Bromcresol

LOT 25AA0501 DATE MANUFACTURER: 2025-05-14

EXP. DATE: 2027-05-15 REF. # GQ-AA-100

**Physical Inspection:** Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

**Performance Characteristics:**

Reagent Albumin @630nm (against water) = 0.149

Aspect Green Liquid Translucid, nonvisible particles, with surfactant

Standard:

Aspect Yellowish Liquid Translucid, nonvisible particles, Value in 4 g/dL

Third Opinion Control: Lyphocheck Assayed Chemistry Control Lot 89740

|                                                 |                           |                 |
|-------------------------------------------------|---------------------------|-----------------|
| Albumin                                         | QCNet (Bio-Rad) Lot 89740 | Results @ 630nm |
| Control 1                                       | n=20 / 4.06 g/dL          | 4.06 (+/- 0.0%) |
| Control 2                                       | n=20 / 2.60 g/dL          | 2.77 (+/- 0.1%) |
| Comparable method with Abbott Architect Systems |                           |                 |

GQ-PC-1 PassCheck Control Lot 25PCK0131/ 25PCK0132

|  |                       |                       |
|--|-----------------------|-----------------------|
|  | Control 1             | Control 2             |
|  | 2.40 g/dL (2.54 g/dL) | 4.70 g/dL (4.78 g/dL) |

Time reader for Abs:                      3 minutes                      5 minutes

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 2.60 g/dL

Storage: 2-25°C

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex<sup>®</sup>