

CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX URIC ACID

LOT: 25UA0501 MFGDT: 2025.05.12

EXP. DATE: 2027.05.15 (YYYY-MM-DD) REF. # GQ-UA-125

BATCH QTY: 90L

Physical Inspection:

Labels: Pass

QR Code: Pass

HPDE Bottle: Pass

Vial: Pass

Box: Pass

Performance Characteristics:

Reagent Uric Acid:

Aspect: Liquid Pink Translucid, non visible particles w/o surfactant

Standard Vol: 1x1ml

Standard Value: 5mg/dL

Blank @505nm (against water) = 0.020

Reagent Stress: Pass

Control Lyphocheck Bio-Rad (LOT 89740)

<https://myeinserts-app.qcnet.com/home>

Liquichek Unassayed Chemistry Control (LOT 92970)

Control 1 (4.54 mg/dL) 4.52 mg/dL

Control 2 (9.83 mg/dL) 9.88 mg/dL

S.D.

<1%

Comparable Method with Abbott Alinity c Analyzer

<1%

Liquichek Urine Chemistry Control (LOT 97440)

Control 1 (4.00 mg/dL) 4.00 mg/dL

Control 2 (9.5 mg/dL) 9.00 mg/dL

PassCheck Universal Assayed Serum Chemistry Control (Lot 25PCK0131, 25PCK0132)

Control 1 (4.30 mg/dL) 4.20 mg/dL

Control 2 (8.50 mg/dL) 8.30 mg/dL

PassCheck Calibrator 24CAL0701

Value: 4.31

Conclusion: Released

