

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Dyazo Total Bilirubin

LOT 25BT0503 DATE MANUFACTURER: 2025-05-21

EXP. DATE: 2027-05-01 REF. # GQ -BT-100

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Total Dyazo Bilirubin

Aspect Liquid Translucid, nonvisible particles, with surfactant

Reagent Nitrite:

Aspect Liquid Translucid, nonvisible particles, with surfactant (Lot 25BDT0501)

Reagent
Abs Work Reagent

Values @ 550nm
(0.010-0.200)

Results @ 550nm
0.040 (within the first 4hrs)

Comparable with GQ-PC-1 PassCheck Control Lot 25PCK0131 / 25PCK0132

Control 1

1.0 mg/dL (0.93)

Control 2

6.20mg/dL (7.10)

Time reader for Abs:

5 minutes


5 minutes (PASS)

Factor for this Lot: 32.25

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 0.93

Storage: 2-25°C

Liable Person:

A circular seal for "BIALEX LAB" with "QUALITY ASSURANCE" around the bottom. In the center, there is a stylized "QC" and three stars. A handwritten signature is written over the seal.

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex[®]