



Passcheck -Microalbumin Calibrator for semiautomatic instrument, Ref. GQ-CMAL-1

Use: IVD

Intended of Use

PassCheck Microalbumin urine calibrator is intended for use in semi-automatic instruments procedures of Bialex's reagent.

Preparation

vial is ready to use, vial x 1mL.

Shelf life:

The vial closed and refrigerated from 2 to 8°C is estable until expiration date indicates in the label.

After opened it is stable for 30 days.

No freezer your vial.

Values:

Updates will be available at www.bialexlab.com/IFU for English version and www.bialexlab.com/insertos for Spanish version

Precaution

All specimens using in the PassCheck Control or Calibrator should be handled in accordance with good laboratory practices using appropriate precautions as described in the CDC/NIH Manual, "Biosafety in Microbiological and Biomedical Laboratories! 2nd edition. 1988 HHS Publication No (CDC)88-8395

Limitations

Due to its easy cross-contamination potential, please handle with care.

Procedures

Use a vortex previously to use, for best homogenization. After each use promptly replace the stopper and return immediately to 2 to 8°C



LOT: 22MAL0501, Expired 2024.05.30

Serum analytical	Value
Microalbumin	3.43

units: mg/dL

(*) You can find update information:

www.bialexlab.com/apps (US Only)

www.bialexlab.com/insertos (Latam Countries)