

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Albumin Bromcresol

LOT 25AA0901 DATE MANUFACTURER: 2025-09-01 (YYYY-MM-DD)
EXP. DATE: 2027-08-30 REF. # GQ-AA-100

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Albumin @630nm (against water) = 0.133

Aspect Green Liquid Translucid, nonvisible particles, with surfactant

Standard:

Aspect Yellowish Liquid Translucid, nonvisible particles, Value in 5 g/dL

Third Opinion Control: Lyphocheck Assayed Chemistry Control Lot 89740

Albumin	QCNet (Bio-Rad) Lot 89740	Results @ 630nm
Control 1	n=20 / 4.32 g/dL	4.31 (+/- 0.1%)
Control 2	n=20 / 2.82 g/dL	2.76 (+/- 0.1%)

Comparable method with Roche / Hitachi Cobas C Systems

GQ-PC-1 PassCheck Control Lot 25PCK0141/ 25PCK0142

	Control 1	Control 2
	2.98 g/dL (3.00 g/dL)	4.50 g/dL (4.50 g/dL)

Time reader for Abs:	3 minutes	5 minutes
----------------------	-----------	-----------

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 3.014 g/dL

Storage: 2-25°C

Liable Person:



Jean Tanaka

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex[®]