

CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX MICROALBUMIN

LOT: 25MAU0901

MFGDT: 2025.08.08(YY/MM/DD)

EXP. DATE: 2027.07.30

REF: GQ-MAU-3010

BATCH: 100 KITS

Physical Inspection:

Label: Pass

QR Code: Pass

HPDE Bottle: Pass

Vial: NRD

Box: Pass

Performance Characteristics:

Reagent 1: Tris Buffer pH 7.4 at 25°C +/-0.1

Aspect: Liquid Translucid, non-visible particles with surfactant

Reagent 2: Substrate

Aspect: Liquid Yellow, non-visible particles w/o surfactant

Accelerated Stress: PASS

BIO-RAD LIQUICHEK URINE CONTROL, Lot 97450

Level 1 = 4.54 mg/dL (5.60) - PASS

Level 2 = 9.50 mg/dL (9.60) – PASS

PASSCHECK URINE CONTROL, Lot 24PMAU10901 & 24PMAU20901

Level 1 = 4.90 mg/dL (5.45) - PASS

Level 2 = 9.40 mg/dL (9.60) - PASS

Optical path: 1cm

Results within the target range

Repeatability CV<4%

CONCLUSION: REALESE FOR SALE

