

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Albumin Bromcresol

LOT 25AA0721 DATE MANUFACTURER: 2025-07-01

EXP. DATE: 2027-07-01(YYYY.MM.DD) REF. # GQ-AA-100

Physical Inspection: Labels: Acceptable | QR Code: Acceptable| HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Albumin @630nm (against water) = 0.152

Aspect Green Liquid Translucid, nonvisible particles, with surfactant

Standard:

Aspect Yellowish Liquid Translucid, nonvisible particles, Value in 4.5 g/dL

(Lot 25AA0701)

Third Opinion Control: Lyphocheck Assayed Chemistry Control Lot 89740

Albumin	QCNet (Bio-Rad) Lot 89740	Results @ 630nm
Control 1	n=20 / 4.06 g/dL	4.06 (+/- 0.0%)
Control 2	n=20 / 2.60 g/dL	2.77 (+/- 0.1%)
Comparable method with Abbott Architect Systems		

GQ-PC-1 PassCheck Control Lot 25PCK0131/ 25PCK0132

	Control 1	Control 2
	2.40 g/dL (2.64 g/dL)	4.67 g/dL (4.78 g/dL)

Time reader for Abs: 3 minutes 5 minutes

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 3.04 g/dL

Storage: 2-25°C

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex[®]