

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Albumin Bromcresol

LOT 25AA0902 DATE MANUFACTURER: 2025-09-23 (YYYY-MM-DD)
EXP. DATE: 2027-09-10 REF. # GQ-AA-100

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Albumin @630nm (against water) = 0.141

Aspect Green Liquid Translucid, nonvisible particles, with surfactant

Standard:

Aspect Yellowish Liquid Translucid, nonvisible particles, Value in 5 g/dL

Third Opinion Control: Lyphocheck Assayed Chemistry Control Lot 89740

Albumin	QCNet (Bio-Rad) Lot 89740	Results @ 630nm
Control 1	n=20 / 4.32 g/dL	4.36 (+/- 0.1%)
Control 2	n=20 / 2.80 g/dL	2.89 (+/- 0.1%)

Comparable results with Abaxis-Piccolo

GQ-PC-1 PassCheck Control Lot 25PCK0141/ 25PCK0142

	Control 1	Control 2
	2.90 g/dL (3.00 g/dL)	4.55 g/dL (4.38 g/dL)

Time reader for Abs:	3 minutes	5 minutes
----------------------	-----------	-----------

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 3.14 g/dL

Storage: 2-25°C

Liable Person:



Jean Tanaka