



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Creatinine

LOT 25CRE0702 **DATE MANUFACTURER:** 2025.07.09 (YYYY-MM-DD)

EXP. DATE: 2027-07-01 **REF. #** GQ-CRE-125

QTY: 700KITS

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable

| Box: Acceptable

Performance Characteristics:

Reagent Creatinine:

Aspect Liquid Translucid Yellow, non-visible particles, with surfactant

Standard: Aspect Liquid Translucid, nonvisible particles, w/o surfactant, concentration **5 mg/dL**

Reagent	▲ Kinetic @37°C	Results @505nm
Reagent Creatinine	Vx 105/1min	0.176
Creatinine	QCNet (Bio-Rad) Lot 89740	Results @ 505nm
Control 2	n=20 / 6.15 mg/dL	6.15 (+/- 0.1%)
Control 1	n=20 / 2.47 mg/dL	2.50 (+/- 0.1%)
Comparable method with Abbott Infinity		

PassCheck Control Level 1 and 2 - Lot 25PCK0130 (Ref: GQ-PC1-1, GQ-PC2-1)

Control 1 0.96 mg/dL (0.96 mg/dL)

Control 2 4.50 mg/dL (4.60 mg/dL)

PassCheck CAL-Lot 24CAL0701

Value: 0.94

Storage: 2-8°C

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®