

CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX LDH

LOT: 25LDH1001

EXP. DATE: 2027.08.30 (YY/MM/DD)

BATCH QTY: 110 KITS

MFGDT: 2027.10.02

REF.: GQ-LDH-125

Physical Inspection:

Labels: Pass

QR Code: Pass

HPDE Bottle: Pass

Vial: Pass

Box: Pass

Performance Characteristics:

BUFFER:

Aspect: Liquid Weak Yellowish Translucid, non visible particles with surfactant

pH: 7.55 at 25°C

Reagent Stress: Pass

SUBSTRATE:

Aspect: Translucid

Reagent Stress: Pass

Work Reagent: 2.5 Abs (1.33)- Pass

Substrate: 1.680 - Pass

Using: 120s read time, 60s interval reader, 20s lag time

PassCheck Universal Assayed Multicontrol (LOT 25PCK0141/2)

Control 1 (206.21U/L)

220.00 U/L

Control 2 (570.81U/L)

550.00 U/L

Programming for automatic biochemistry instruments: 120 s (Read Time) / 60 s (Interval Time) / 20 s (Lag time),
F=196.21

Programming for semi-automatic biochemistry instruments: 120 s (Read Time) / 60 s (Interval Time) / 20 s (Lag time)
F=7567.56

Conclusion: Released

