

**PRODUCT NAME: Bialex LDL**

LOT 25LDL1101 DATE MANUFACTURER: 2025.10.20

EXP. DATE: 2027-10-01 REF. # GQ-LDL-3010

Qty: 190 Kits

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable

| Box: Acceptable

Method: Direct

Performance Characteristics:**Reagent 1:**

Aspect Liquid Translucid Light brown, non-visible particles, w/o surfactant

Reagent 2:

Aspect Translucid, non-visible particles.

Color reaction: Purple

	QCNet (Bio-Rad)	Results @ 600nm
Control 1	153.0 mg/dL	151.80 mg/dL (n=20)
Control 2	52.3 mg/dL	53.00 mg/dL (n=20)

Comparable method with Beckman Coulter AU System

PassCheck -(Lot 24PCK0141 & 25PCK0142)

Control 1	56.00 (52 mg/dL)
Control 2	130.00 (133 mg/dL)

Serum Abnormal >0.170 Pass

Serum Normal <0.090 Pass

PassCheck CAL Ref: GQ-CAL-5 (LOT:24CAL0701)

Value:61 mg/dL

Conclusion: Realease Batch**Liabe Person:**