



CERTIFICATE OF ANALYSIS

Product Name: Chlorine (Cl-) Test Kit (Mercury Thiocyanate)		LOT 51112141H		REF: 709141	
Package	R:20mlx8	MFG by:	BIOBASE BIODUSTRY (SHANDONG) Co.	MFG Date	2025.11.13
Exp. Date	2026.11.12	Appearance and Character	Yellow clear	Reagent blank	0.0202 (at 436nm)
Accuracy	Within the target range	Linear	130mmol/L	Repeatability	<8%
Conclusion:	Pass				

