

CERTIFICATE OF ANALYSIS

Product Name: UREA (UREA) TEST KIT (UREASE-GLDH)		LOT:51114024H		REF:71324	
Package	R1:48mLx 2 R2:16mLx2	MFG by:	BIOBASE BIODUSTRY (SHANDONG) Co.	MFG Date	2025.11.17
Exp. Date	2026.11.16	Appearance and Character	Light yellow clear liquid (R1 / R2)	Reagent blank	1.576(at 340nm)
Accuracy	Within the target range	Linear	35.7mmol/L	Repeatability	<3.54%
Conclusion:	Pass				

