

CERTIFICATE OF ANALYSIS

Name	HbA1c Control 1 and 2	Lot No.(Product) Exp. Date:	50923162L 2026.09.16
specifications	1mL×2	Date of Manufacture	2025.9.30
Lot No. (Calibrator)	50912162L	Lot No.(Reagent)	50827097F
Appearance	red	Content	1.0mL

Sampling volume	S: 25μL; R1: 160μL; R2: 70μL			Analytical method	end-point method
Primarily	32-33	Secondary	12-13		

Accuracy						
Target value	Range	1	2	3	Average value	
5.2	4.7-5.7	5.4	5.3	5.1	5.3	
10.1	9.1-11.1	9.8	10.1	9.8	9.9	
Box	Results	Repeated test results				
1	5.3	5.2	5.0	5.4	5.04	
2	5.0	5.1	5.4	5.4	CV between the bottles	1.04%
3	5.3	5.0	5.4	5.1		
Box	Results	Repeated test results				
1	10.4	9.8	10.4	10.1	10.10	
2	10.3	10.3	9.8	10.2	CV between the bottles	0.00%
3	10.1	10.4	9.8	10.1		

