



CERTIFICATE OF ANALYSIS

Date: 2025-07-20 (YYYY-MM-DD)

Lot: 25NA0702

Batch: 90 Kits

Property Name: BIALEX Na

Ref: GQ-NA-100

Use: IN VITRO DIAGNOSTIC

Expiration Date: 2027-07-01

Appearance: Color, clarity between parameters

Chemical Stability: Pass

Handling Stress: Pass

Accelerated stability: Pass

Control Reference Values

MEDIA VALUES:

CONTROL 1: 128.50 / CONTROL 2: 150

Unites: mEqL

Result:

LEVEL 1: 130 mEqL / LEVEL 2: 162 mEqL

This lot is ready for release: Ready



J. TANAKA