

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Albumin Bromcresol

LOT 26AA0201 DATE MANUFACTURER: 2026-01-05

EXP. DATE: 2028-01-30(YYYY.MM.DD) REF. # GQ-AA-100

Physical Inspection: Labels: Acceptable | QR Code: Acceptable| HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Albumin @630nm (against water) = 0.161

Aspect Green Liquid Translucid, nonvisible particles, with surfactant

Standard:

Aspect Yellowish Liquid Translucid, nonvisible particles, Value in 5 g/dL,

Lot 26AA0202

Third Opinion Control: Lyphocheck Assayed Chemistry Control Lot 89740

Albumin	QCNet (Bio-Rad) Lot 89740	Results @ 630nm
Control 1	n=20 / 4.06 g/dL	4.1 (+/- 0.5%)
Control 2	n=20 / 2.60 g/dL	2.70 (+/- 0.1%)
Comparable method with Abbott Architect Systems		

GQ-PC-1 PassCheck Control Lot 25PCK0161/ 25PCK0162

	Control 1	Control 2
	2.40 g/dL (2.64 g/dL)	4.67 g/dL (4.78 g/dL)

Time reader for Abs: 3 minutes 5 minutes

PassCheck CAL Lot: 24CAL0701 (Ref: GQ-CAL-5) - Value: 3.04 g/dL

Storage: 2-25°C

Liable Person:

