



Certificate of Analysis

Date MNFG: 2026.03.10
Lot: 26K0301
Property Name: BIALEX K

Batch: 400 Kits

Reference: GQ-K-100
Use: IN VITRO DIAGNOSTIC
Expiration Date: 2028-03-03 (YYYY-MM-DD)

Characteristics:
REAGENT AND STANDARD BETWEEN RANGE CHARACTERISTIC
Control Reference:

RANGE VALUES
CONTROL 1: 4.6
CONTROL 2 6.3
Units: mEq/L
St. Value: 4mEq/L

Result: LEVEL 1 3.95 / LEVEL 2 6.70 mEq/L

Conclusion: Release for sale

A handwritten signature in blue ink is written over a circular blue stamp. The stamp contains the text "BIALEX", "QC", "DATE", and "QUALITY ASSURANCE".