

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Triglycerides

LOT 26TT0301 DATE MANUFACTURER: 2026-02-20

EXP. DATE: 2028-02-10 (YYYY-MM-DD) REF. # GQ-TT-100, GQ-TT-125, GQ-TT-250, GQ-TT-2X125, GQ-TT-4X125

Qty Manufacture: 100L

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable

Performance Characteristics:

Reagent Triglycerides:

Aspect Liquid Clear Pink (Source Candide Rugose) to pink,
visible particles, with surfactant 1%

Interference blocker

Standard:

Include in the kit – Use Standard of Triglycerides Bialex (GQ-TT-1), (25TT1101)

Concentration 200mg/dL

Reagent

Reagent Triglycerides Abs:

Reagent Triglycerides

Control 1

Control 2

Comparable method with Abbott Alinity c Analyzer (continued)

Values @ 505nm

QCNet (Bio-Rad) Lot 89740

n=20 / 153 mg/dL

n=20 / 77.8 mg/dL

Results @ 505nm

0.008

Results @ 505nm

152.0 mg/dL

76.0 mg/dL

PassCheck -Lot 26PCK0171, 26PCK0172 (REF: GQ-PC1-1, GQ-PC2-1)

Control 1

89.00 mg/dL

Control2

203.48 mg/dL

PassCheck CAL-Lot 26CAL0101 (REF: GQ-CAL-5)

Value 112 mg/dL

Linearity: up 1100 mg/dL (PASS)

Stress stability: 3 days (78 hrs) (PASS)

Recommendations: Ultra sensible to cross contaminations, keep out of direct light.

Storage 2-8°C

Conclusion: Release for sale

Liable Person: Jean Tankaka

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex®