



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Creatinine

LOT 26CRE0401 **DATE MANUFACTURER:** 2026.04.14 (YYYY-MM-DD)

EXP. DATE: 2028-04-06 **REF. #** GQ-CRE-125

QTY: 160KITS

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable

Performance Characteristics:

Reagent Creatinine:

Aspect Liquid Translucid Yellow, non-visible particles, with surfactant

Standard: Aspect Liquid Translucid, nonvisible particles, w/o surfactant, **5 mg/dL**

Reagent (LOT 26CRE0401)	▲ Kinetic @37°C	Results @505nm
Reagent Creatinine	Vx 107/1min	0.191

Creatinine	QCNet (Bio-Rad) Lot 89740	Results @ 505nm
Control 2	n=20 / 6.15 mg/dL	6.10 (+/- 0.1%)
Control 1	n=20 / 2.47 mg/dL	2.39 (+ 0.1%)
Comparable method with Abbott Infinity		

PassCheck Control Level 1 and 2 - Lot 26PCK0170 (Ref: GQ-PC1-1, GQ-PC2-1)

Control 1 1.00 mg/dL (1.13 mg/dL)

Control 2 4.70 mg/dL (4.30 mg/dL)

PassCheck CAL-Lot 26CAL0101

Value: 1.10

Storage: 2-8°C

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®