

**PRODUCT NAME: Bialex LDL**

LOT 26LDL0302 DATE MANUFACTURER: 2026.03.12

EXP. DATE: 2028-02-10 REF. # GQ-LDL-3010

Qty: 30 Kits

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable

| Box: Acceptable

Method: Direct

Performance Characteristics:**Reagent 1:**

Aspect Liquid Translucid Light brown, non-visible particles, w/o surfactant

Reagent 2:

Aspect Translucid, non-visible particles.

Color reaction: Purple

	QCNet (Bio-Rad)	Results @ 600nm
Control 1	153.0 mg/dL	150.00 mg/dL (n=20)
Control 2	52.3 mg/dL	49.00 mg/dL (n=20)

Comparable method with Beckman Coulter AU System

PassCheck -(Lot 26PCK0171 & 26PCK0172)

Control 1	58.00 (59.33 mg/dL)
Control 2	127.00 (126.00 mg/dL)

Serum Abnormal >0.170 Pass

Serum Normal <0.090 Pass

PassCheck CAL Ref: GQ-CAL-5 (LOT:26CAL0101)

Value:56.68 mg/dL

Conclusion: Realease Batch**Liabe Person:**

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.**TollFree: 833-418-2116 email: info@bialexlab.com**