

CERTIFICATE OF ANALYSIS

Product: PassCheck Urine Calibrator MicroAlbumin

Reference: GQ-CMAL-1

Lot Number: 26CMAL801

Manufacturing Date: 2026-04-30

Expiration Date (Unopened): 2028-01-10

Open Vial Stability (2-8°C): 10 days

Reference Material Correlation:

Internal traceability and accuracy were verified by side-by-side comparison with

Bio-Rad Lyphochek[®] Quantitative Urine Control, Lot #1007110.

Validation Results:

- **Correlation Analysis:** Analytical runs using Lot 25MAU1102 demonstrated full alignment with Bio-Rad Lot #1007110 and 26CMAL801
- **Result: Equivalency confirmed.** The recovered values for Microalbumin in the current lot showed no statistically significant bias ($p > 0.05$) compared to the Bio-Rad reference material.

