



Certificate of Analysis

Product Name	Sodium (Na) Test Kit (Enzymatic)			Material Code	71339
Lot No.	60127151T	Package	R1:60ml×2 R2:20ml×2	Manufacture Date	2026.2.16
Lot Size	R1:60ml×2 R2:20ml×2	Market Area	All countries	Expiry Date	2028.2.15
Sample Amount	1	Sampling Date	2026.2.16	Analysis Date	2026.2.16
Inspection Basis	Inspection procedures				
Test Items	Quality Standard			Result	Comment
Appearance and Character	Reagent componets are complete.Labels are clear and correct.			Reagent componets are complete.Labels are clear and correct.	Qualified
	R1 R1 is colorless clear liquid			R1 R1 is colorless clear liquid	
	R2 R2 is colorless clarifying liquid			R2 R2 is colorless liquid	
Volume	R1:60ml×2 R2:20ml×2 R1:60ml×2 R2:20ml×2			R1 60.3 ml R2 20.2ml	Qualified
Blank Absorbance	405nm 1cm 0.5 Reagent blank absorbance ≤0.5,(405nm,1cm optical path)			0.2339	Qualified
Accuracy	Whether the quality control results are within the target range			Yes	Qualified
Linear Range	[80,180] mmol/L range: [80,180]mmol/L, $r \geq 0.990$			0.9991	Qualified
Reproducibility	CV 5% repeatability CV ≤5%			1.52%	Qualified
Conclusion	Pass				
Remarks					
QC		Checked By		Approved By	
Signature / Date	2026.2.16	Signature / Date	2026.2.16	Signature / Date	2026.2.16