

CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX URIC ACID

LOT: 26UA0701 MFGDT: 2026.06.26

EXP. DATE: 2028.06.20 (YYYY-MM-DD)

BATCH QTY: 101L

REF. GQ-UA-125, GQ-UA-250

Physical Inspection:

Labels: Pass

QR Code: Pass

HPDE Bottle: Pass

Vial: Pass

Box: Pass

Performance Characteristics:

Reagent Uric Acid:

Aspect: Liquid Pink Translucid, non visible particles w/o surfactant

Standard Vol: 1x1ml

Standard Value: 5 mg/dL

Blank @505nm (against water) = 0.010

Reagent Stress: Pass

Control Lyphocheck Bio-Rad (LOT 89740)

<https://myeinserts-app.qcnet.com/home>

Liquichek Unassayed Chemistry Control (LOT 93030)

Control 1 (4.00 mg/dL)

4.10 mg/dL

S.D.

<1%

Control 2 (9.50mg/dL)

9.50mg/dL

<1%

Liquicheck Urine Chemistry Control (LOT 97470)

Control 1 (9.87 mg/dL)

9.40 mg/dL

Control 2 (15.70 mg/dL)

15.60 mg/dL

Linearity: up 30 mg/dL

PassCheck Universal Assayed Serum Chemistry Control (Lot 26PCK0171, 26PCK0172)

Control 1 (4.50 mg/dL)

4.40 mg/dL

Control 2 (8.70 mg/dL)

8.30 mg/dL

PassCheck Calibrator 26CAL0101

Value: 4.5

Conclusion: Released

