



CERTIFICATE OF ANALYSIS

Date: 2026-04-30 (YYYY-MM-DD)
Lot: 26Na0601
Batch: 130 Kits

Property Name: BIALEX Na
Ref: GQ-NA-100
Use: IN VITRO DIAGNOSTIC
Expiration Date: 2028-03-30

Appearance: Color, clarity between parameters

Chemical Stability: Pass

Handling Stress: Pass

Accelerated stability: Pass

Control Reference Values

MEDIA VALUES:

CONTROL 1: 130.00 / CONTROL 2: 150

Unites: mEqL

Result:

LEVEL 1: 130 mEqL / LEVEL 2: 162 mEqL

This lot is ready for release: Ready

