



Certificate of Analysis

Product Name: Bialex Ferrozine

Ref: GQ-UIBC-100

LOT: 26UBIC0601 EXP. DATE: 2028.04.10 (YYYY-MM-DD)

We conducted an analysis of 20 samples and compared the results obtained with those produced using a commercial reagent. The findings are detailed below:

The standard deviation observed across the runs was found to be less than 1%, indicating a high level of precision in the measurements.

The assay was conducted using special serum to determine linearity, and the results indicated a concentration of 500 µg/dL.

The following values were obtained using the Bio-Rad Control Lot 89740

Level I 237 µg/dL (236.00 µg/dL)

Level II 74.3 µg/dL (76.13 µg/dL)

The batch underwent stress testing and successfully passed all assessments

Conclusions:

Release for sale

