

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Total Protein

LOT 26TPB0602 DOM: 2026-06-25

EXP. DATE: 2028-06-20 (YYYY-MM-DD), REF. # GQ-TPB-125

Qty Manufacture: 60 Liters

Physical Inspection:

Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:

Reagent Total Protein:

Aspect Liquid Translucid-blue clear, non-visible particles, w/o surfactant

Standard:

Include in the kit – Use Standard of Bialex Total Protein (GQ-TPB-1),

Aspect Yellow

Concentration 4.5 g/dL

Reagent

Reagent Total Protein

Reagent Total Protein

Control 1

Control 2

Comparable method with Abaxis Piccolo.

Values @ 546nm

(0.060-0.900)

QCNet (Bio-Rad) Lot 89740

n=20 / 6.70 g/dL

n=20 / 4.30 g/dL

Results @ 546nm

0.067

Results @ 540nm

6.50 g/dL

4.10 g/dL

GQ-PC-1 PassCheck Multicontrol - LOT 26PCK0171 / 26PCK0172

Control 1

3.84 g/dL

(3.90 g/dL)

Control 2

7.20 g/dL

(7.00g/dL)

Cal Value (26CAL0101) =3.51 g/dL

Time reader for Abs: 45m

Pass

Storage: 2-8°C

Conclusion: Release for sale

Liabe Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex®