



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Creatinine

LOT 26CRE0601 **DATE MANUFACTURER:** 2026.06.25 (YYYY-MM-DD)

EXP. DATE: 2028-06-20 **REF. #** GQ-CRE-125, GQ-CRE-250

QTY: 600KITS (GQ-CRE-125)

QTY: 90 KITS (GQ-CRE-250)

Physical Inspection: Labels: Acceptable | QR Code: Acceptable | HPDE Bottle: Acceptable
| Box: Acceptable

Performance Characteristics:

Reagent Creatinine:

Aspect Liquid Translucid Yellow, non-visible particles, with surfactant

Standard: Aspect Liquid Translucid, nonvisible particles, w/o surfactant, **5 mg/dL**

Reagent (LOT 26CRE0601)
Reagent Creatinine

▲ Kinetic @37°C
Vx 107/1min

Results @505nm
0.189

Creatinine

QCNet (Bio-Rad) Lot 89740

Results @ 505nm

Control 2

n=20 / 6.15 mg/dL

6.00 (+/- 0.1%)

Control 1

n=20 / 2.47 mg/dL

2.6 (+ 0.6%)

Comparable method with Abbott Infinity

PassCheck Control Level 1 and 2 - Lot 26PCK0170 (Ref: GQ-PC1-1, GQ-PC2-1)

Control 1

1.00 mg/dL (0.99 mg/dL)

Control 2

4.70 mg/dL (4.70 mg/dL)

PassCheck CAL-Lot 26CAL0101

Value: 1.10

Storage: 2-8°C

Liabe Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®