Bialex®

# CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX MICROALBUMIN
LOT: 26MAU0701 MFGDT: 2026.07.11 (YY/MM/DD)
EXP. DATE: 2028.02.10 REF: GQ-MAU-3010
BATCH: 90 KITS

Physical Inspection:
Label: Pass QR Code: Pass HPDE Bottle: Pass Vial: NRD Box: Pass

Performance Characteristics:
Reagent 1: Tris Buffer pH 7.4 at 25ºC +/-0.1
Aspect: Liquid Translucid, non-visible particles with surfactant
Reagent 2: Substrate
Aspect: Liquid Yellow, non-visible particles w/o surfactant
Accelerated Stress: PASS

BIO-RAD LIQUICHEK URINE CONTROL, Lot 97450
Level 1 = 4.54 mg/dL (5.00) - PASS
Level 2 = 9.50 mg/dL (9.70) – PASS
PASSCHECK URINE CONTROL, Lot 26CMAU0701/2
Level 1 = 4.00 mg/dL (4.07) - PASS
Level 2 = 10.00 mg/dL (9.90) - PASS

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.
TollFree: 833-418-2116 email: info@bialexlab.com