Bialex®

CERTIFICATE OF ANALYSIS

Date: 2026-03-30 (YYYY-MM-DD)
Lot: 26Na0301
Batch: 28 Kits

Property Name: BIALEX Na
Ref: GQ-NA-100
Use: IN VITRO DIAGNOSTIC
Expiration Date: 2028-01-20

Appearance: Color, clarity between parameters

Chemical Stability: Pass

Handling Stress: Pass

Accelerated stability: Pass

Control Reference Values
MEDIA VALUES:
CONTROL 1: 130.00 / CONTROL 2: 150
Unites: mEqL
Result:
LEVEL 1: 130 mEqL / LEVEL 2: 162 mEqL
This lot is ready for release: Ready

BIALEX LAB QC DATE QUALITY ASSURANCE

J.TANAKA

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex®

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.
TollFree: 833-418-2116 email: info@bialexlab.com