Bialex®

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Creatinine

LOT 26CRE0802 **DATE MANUFACTURER**: 2026.08.04 (YYYY-MM-DD)
EXP. DATE: 2028-07-10 **REF.** # GQ-CRE-125, GQ-CRE-250
QTY: 100 KITS (GQ-CRE-125)

Physical Inspection: Labels: Acceptable | QR Code: Acceptable| HPDE Bottle: Acceptable | Box: Acceptable
Performance Characteristics:
Reagent Creatinine:
Aspect Liquid Translucid Yellow, non-visible particles, with surfactant
Standard: Aspect Liquid Translucid, nonvisible particles, w/o surfactant, **5 mg/dL**

Reagent (LOT 26CRE0802)	▲Kinetic @37ºC	Results @505nm
Reagent Creatinine	Vx 104/1min	0.200

Creatinine	QCNet (Bio-Rad) Lot 89740	Results @ 505nm
Control 2	n=20 / 6.15 mg/dL	6.00 (+/- 0.1%)
Control 1	n=20 / 2.47 mg/dL	2.6 (+ 0.6%)

Comparable method with Abbott Infinity

PassCheck Control Level 1 and 2 - Lot 26PCK0170 (Ref: GQ-PC1-1, GQ-PC2-1)

Control 1	1.00 mg/dL (1.07 mg/dL)
Control 2	4.70 mg/dL (4.67 mg/dL)

PassCheck CAL-Lot 26CAL0101
Value:1.07

Storage: 2-8ºC

Liable Person:



Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum.
All Rights Reserved by Bialex®

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.
TollFree: 833-418-2116 email: info@bialexlab.com