Bialex®

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Bialex Triglycerides
LOT 26TT0702 DATE MANUFACTURER: 2026-07-07 TO 2026-07-17
EXP. DATE: 2028-07-01 (YYYY-MM-DD) REF. # GQ-TT-100, GQ-TT-125,GQ-TT-250,GQ-TT-2X125,GQ-TT-4X125

Qty Manufacture: 300L

Physical Inspection:
Labels: Acceptable | QR Code: Acceptable| HPDE Bottle: Acceptable | Box: Acceptable

Performance Characteristics:
Reagent Triglycerides:
Aspect Liquid Clear Pink (Source Candide Rugose) to pink, visible particles, with surfactant 1%
Interference blocker
Standard:
Include in the kit – Use Standard of Triglycerides Bialex (GQ-TT-1), Concentration 200mg/dL (LOT 26TT0701)

Reagent	**Values @ 505nm**	**Results @ 505nm**
Reagent Triglycerides Abs:	----------	0.014
Reagent Triglycerides	QCNet (Bio-Rad) Lot 89740	Results @ 505nm
Control 1	n=20 / 153.0 mg/dL	150.0 mg/dL
Control 2	n=20 / 77.8 mg/dL	81.0 mg/dL

Comparable method with Abbott Alinity c Analyzer (continued)

PassCheck -Lot 26PCK0171, 26PCK0172 (REF: GQ-PC1-1,GQ-PC2-1)

Control 1	Control2
88.72 mg/dL	212.00 mg/dL

PassCheck CAL-Lot 26CAL0101 (REF: GQ-CAL-5)
Value 88.72 mg/dL
Linearity: up 1000 mg/dL (PASS)
Stress stability: 3 days (78 hrs) (PASS)
Recommendations:Ultra sensible to cross contaminations, keep out of direct light.
Storage2-8ºC
Conclusion: Release for sale
Liable Person: J.Tankaka

BIALEXLAB QC DATE QUALITY ASSURANCE

Vector must not be considered pathogenic for livestock or avian species; the kit does not contain genes or express antigens of livestock or poultry disease agents. The kit does not contain any animal derived additives, such as albumin or serum. All Rights Reserved by Bialex®

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.
TollFree: 833-418-2116 email: info@bialexlab.com