Bialex®

# CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX URIC ACID
LOT: 26UA0601 MFGDT: 2026.06.01
EXP. DATE: 2028.06.01 (YYYY-MM-DD) REF. GQ-UA-250
BATCH QTY: 30L

Physical Inspection:
Labels: Pass QR Code: Pass HPDE Bottle: Pass Vial: Pass Box: Pass

Performance Characteristics:

Reagent Uric Acid:
Aspect: Liquid Pink Translucid, non visible particles w/o surfactant
Standard Vol: 1x1ml
Standard Value: 6mg/dL
Blank @505nm (against water) = 0.0220
Reagent Stress: Pass
Control Lyphocheck Bio-Rad (LOT 89740)
https://myeinserts-app.qcnet.com/home

| | | S.D. |
|---|---|---|
| Liquichek Unassayed Chemistry Control (LOT 93030) | | |
| Control 1 (4.00 mg/dL) | 4.20 mg/dL | <1% |
| Control 2 (9.50mg/dL) | 9.70mg/dL | <1% |

Liquicheck Urine Chemistry Control (LOT 97470)

| | |
|---|---|
| Control 1 (9.87 mg/dL) | 9.20 mg/dL |
| Control 2 (15.70 mg/dL) | 15.90 mg/dL |

Linearity: up 30 mg/dL

PassCheck Universal Assayed Serum Chemistry Control (Lot 26PCK0171, 26PCK0172)

| | |
|---|---|
| Control 1 (4.50 mg/dL) | 4.30 mg/dL |
| Control 2 (8.70 mg/dL) | 8.90 mg/dL |

PassCheck Calibrator 26CAL0101
Value: 4.4

Conclusion: Released

BIALEX LAB QC DATE QUALITY ASSURANCE

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.
TollFree: 833-418-2116 email: info@bialexlab.com