Bialex®

717 Ohio Ave, Wichita Falls, TX 76301

PassCheck Urine Control for Microalbuminuria, Creatinine and other analytes

FORM 195 Rev No: New

Use: Quality control in clinical urinalysis (In Vitro Diagnostic - IVD) / **Uso:** Control de calidad en uroanálisis clínico (Diagnóstico In Vitro - IVD)

Items: GQ-CMAU1-1 & GQ-CMAU2-1

INTENDED USE

PassCheck Urine Control is a human urine-based control solution specifically designed for the daily monitoring of precision and analytical performance, such as the application of **Westgard multi-rules**. Formulated for clinical chemistry testing across a variety of laboratory platforms, it ensures consistent diagnostic accuracy and procedural integrity.

PREPARATION

- **Ready to use:** No reconstitution required.
- Mix gently by inversion before sampling.

STORAGE AND STABILITY

- Store unopened or opened vials at **2 - 8°C.**
- After opening, the control solution is stable for **24 hours** when tightly capped and stored properly.

PRECAUTIONS

Handle all specimens and controls in accordance with good laboratory practices. Follow universal biohazard precautions as described in the CDC/NIH Manual, *"Biosafety in Microbiological and Biomedical Laboratories"*, 2nd Edition, 1988 (HHS Publication No. CDC 88-8395).

USO PREVISTO

El control de orina PassCheck es una solución de control a base de orina humana diseñada específicamente para el monitoreo diario de la precisión y el desempeño analítico, como la aplicación de las **multireglas de Westgard**. Formulado para pruebas de química clínica en diversas plataformas de laboratorio, garantiza la consistencia y la integridad de los procesos de diagnóstico.

PREPARACIÓN

- **Listo para usar:** No requiere reconstitución.
- Mezcle suavemente por inversión antes de tomar la muestra.

ALMACENAMIENTO Y ESTABILIDAD

- Conserve los viales (abiertos o cerrados) entre **2 y 8°C.**
- Después de abierto, la solución de control permanece estable por **24 horas** si se mantiene correctamente tapada y refrigerada.

PRECAUCIONES

Manipule todas las muestras y controles de acuerdo con las buenas prácticas de laboratorio. Siga las precauciones universales de riesgo biológico descritas en el manual de la CDC/NIH, *"Biosafety in Microbiological and Biomedical Laboratories"*, 2.ª edición, 1988 (Publicación HHS No. CDC 88-8395).