Bialex®

CERTIFICATE OF ANALYSIS

PRODUCT NAME: BIALEX URIC ACID
LOT: 26UA0906 MFGDT: 2026.10.02
EXP. DATE: 2028.10.01 (YYYY-MM-DD) REF. : GQ-UA-250
BATCH QTY: 10L

Physical Inspection:
Labels: Pass QR Code: Pass HPDE Bottle: Pass Vial: Pass Box: Pass

Performance Characteristics:

Reagent Uric Acid:
Aspect: Liquid Pink Translucid, non visible particles w/o surfactant
Standard Vol: 1x1ml
Standard Value: 5 mg/dL
Blank @505nm (against water) = 0.070
Reagent Stress: Pass
Control Lyphocheck Bio-Rad (LOT 89740)
https://myeinserts-app.qcnet.com/home

		S.D.
Liquichek Unassayed Chemistry Control (LOT 93030)		
Control 1 (4.00 mg/dL)	4.19 mg/dL	<1%
Control 2 (9.50mg/dL)	9.10mg/dL	<1%

Liquicheck Urine Chemistry Control (LOT 97490)	
Control 1 (9.87 mg/dL)	9.40 mg/dL
Control 2 (15.70 mg/dL)	15.60 mg/dL

Linearity: up 30 mg/dL

PassCheck Universal Assayed Serum Chemistry Control (Lot 26PCK0171, 26PCK0172)

Control 1 (4.50 mg/dL)	4.40 mg/dL
Control 2 (8.70 mg/dL)	8.30 mg/dL

PassCheck Calibrator 26CAL0101
Value: 4.5

Conclusion: Released

BIALEX LAB QC DATE QUALITY ASSURANCE

717 Ohio Ave. Wichita Falls, TX 76301 - U.S.A.
TollFree: 833-418-2116 email: info@bialexlab.com